



**Report on the Calibration of Three Neutral Density Filters and
One Holmium Filter**

Report No. Photometry/2009/384, 17 February 2009.

Issued by the

Measurement Standards Laboratory of New Zealand

Established under the Measurement Standards Act.1992 and the National Standards Regulations 1976 (with Amendment No. 1, 1992) to provide for uniform measurement of physical quantities throughout New Zealand. All results quoted in this report are directly traceable to the New Zealand national measurement standards held by the Measurement Standards Laboratory of New Zealand, Industrial Research Limited.

JDB *M*

Report on the Calibration of Three Neutral Density Filters and One Holmium Filter

Description and Identification

Three neutral density glass filters and one holmium filter. The filters were mounted in standard sized spectrophotometer sample holders.

The filters were supplied in a container labelled as follows:

Perkin Elmer Standard for UV/Vis
Spectrophotometry B050 - 7805

The filters were labelled on their holders and the glass as follows:

Holder	Glass
G1	G1 1886
G2	G2 1886
G3	G3 1506
H	Ho 1826

Client

New Zealand Scientific Ltd, Kyocera Building, 1 - 3 Parkhead Place, Albany 0632, Auckland

Date of Calibration

4 to 9 February 2009.

Objective

- Neutral Density Filters

To calibrate the optical density of the three neutral density filters at the following wavelengths: 440 nm, 546.1 nm, and 635 nm.

This report may be reproduced in full, but not in part, without the written consent of the Chief Metrologist, Measurement Standards Laboratory of New Zealand

704
h

- Holmium Filter

To calibrate the wavelengths of the absorption peaks in the filter at the following nominal wavelengths: 279 nm, 361 nm, 460 nm and 536 nm.

Technical Procedure

MSLT.O.006.007

Method of Calibration

Before calibration the filters were removed from their holders and cleaned using methanol and lens tissue.

- Neutral Density Filters

Transmittance measurements were made on the filters using a high accuracy spectrophotometer system. Measurements were made on the central 4 mm x 10 mm area of the filters using collimated light at the required wavelengths and with a bandwidth of 1.0 nm. The optical densities (OD) of the filters were calculated from the transmittance values (T) using the formula:

$$OD = -\log_{10}T$$

- Holmium Filter

The spectral transmittance at wavelengths around the absorption peaks of the filter were measured on a high accuracy spectrophotometer system using collimated light with a bandwidth of 0.2 nm. Measurements were made at wavelength intervals of 0.1 nm through the peaks to find the wavelength of minimum transmittance.

Conditions

These calibration were carried out at a room temperature of 20 °C ± 0.5 °C.

Results

- Neutral Density Filters

The following table lists the neutral density filters, the wavelength of calibration and the corresponding optical density.

Wavelength nm	Filter Identifier		
	G1	G2	G3
440	0.338	0.989	0.499
546.1	0.307	0.946	0.467
635	0.332	0.941	0.483

This report may be reproduced in full, but not in part, without the written consent of the Chief Metrologist, Measurement Standards Laboratory of New Zealand

704 M n

- Holmium Filter

The following table lists the wavelengths of minimum transmittance of the absorption lines for the holmium filter.

279.3 nm
360.8 nm
460.2 nm
536.3 nm

Uncertainty

The uncertainty in the optical density values at the 95% confidence level and with a coverage factor of 2.5 is 0.002. The uncertainty at the 95% confidence level in the wavelength is 0.3 nm with a coverage factor of 2.2. (See ISO guide to the Expression of Uncertainty in Measurement 1995, for an explanation of terms.)



J D Hamlin
Technical Officer



K M Nield
Research Scientist



P Saunders
for Chief Metrologist